



CERTIFICATE OF ANALYSIS

ISO 9001-2015 CERTIFIED

Product Name : ALUMINIUM AAS STANDARD SOLUTION 1000 mg/L Al IN DILUTED HCl traceable to NIST Expiry date: 12 months from the date of opening

Batch No. : Sample COA Release Date : 18-Mar-2025

Mol. Formula : Mol. Weight :

Code No. : A200H CAS No. : NA

Mfg. Date : Mar-2025 Exp. Date : Feb-2029

HAZ. / P.G. : 8 / III UN No. : 3264

Sr.	Tests	Specifications	Results
1	Appearance	A clear colourless liquid	A clear colourless liquid
2	Amount of substance concentration c(Al)	1000 mg/L +/-0.2%	999.38mg/L
3	Matrix	2% HCl	Complies

CONCLUSION -This above product complies as per the specifications of LOBA CHEMIE PVT. LTD.

Kavita
Chemist

Javed Tamboli
QC Manager

LOBA CHEMIE PVT. LTD.

Mfg. Add.: LOBA Chemie Pvt. Ltd. Plot No. D-22, MIDC Tarapur, Boisar, Tal. & Dist.-Palghar, Maharashtra-401506
Tel: 91.02525-663630/39/34

HO Add.: Loba Chemie Pvt. Ltd., Jehangir Villa, 107, Wodehouse Road., Colaba, Mumbai, Maharashtra – 400005
Customer Care No: +9122 66636663/Email:info@lobachemie.com www.lobachemie.com